

**Columbus Consolidated Government  
Council Meeting Agenda Item**

<b>TO:</b>	Mayor and Councilors
<b>AGENDA SUBJECT:</b>	Stryker Lifepak 15 Monitor Defibrillator and Accessories for Fire & EMS Department
<b>INITIATED BY:</b>	Finance Department

---

It is requested that Council approve the purchase of one (1) Stryker Lifepak 15 Monitor Defibrillator and accessories from Stryker Medical (Chicago, IL) in the amount of \$45,346.29.

The Columbus Fire & EMS Department has standardized to using Stryker (formerly Physio Control) cardiac monitors and automatic external defibrillators on all apparatus. The Stryker LifePak 15 is a full Cardiac Monitor utilized by Paramedics on ambulances and engines. It can run 3-lead ECG, 12-lead ECG, allow for cardiac pacing, provide manual defibrillation, monitor end-tidal CO<sub>2</sub>, monitor blood pressure and measure SPO<sub>2</sub>. This unit can only be used by personnel trained to the paramedic level and is the leading equipment used when providing advanced cardiac care.

Stryker/Physio Control is the manufacturer of the Lifepak 15. Consequently, to maintain compatibility with the remaining units that are in-service and to receive appropriate trade-in credit for older units, the vendor is considered the only known source for this purchase per the Procurement Ordinance, Article 3-114.

Funds are budgeted in the FY24 Budget: General Fund - Fire & EMS – Fire & EMS Operations – Capital Expenditure Over \$5,000; 0101 – 410 – 2100 – FOPR – 7761.

**A RESOLUTION**

**NO.**\_\_\_\_\_

**A RESOLUTION AUTHORIZING THE PURCHASE OF ONE (1) STRYKER LIFEPAK 15 MONITOR DEFIBRILLATOR AND ACCESSORIES FROM STRYKER MEDICAL (CHICAGO, IL) IN THE AMOUNT OF \$45,346.29.**

**WHEREAS**, the Columbus Fire & EMS Department has standardized to using Stryker (formerly Physio Control) cardiac monitors and automatic external defibrillators on all apparatus. The Stryker LifePak 15 is a full Cardiac Monitor utilized by Paramedics on ambulances and engines. It can run 3-lead ECG, 12-lead ECG, allow for cardiac pacing, provide manual defibrillation, monitor end-tidal CO2, monitor blood pressure and measure SPO2. This unit can only be used by personnel trained to the paramedic level and is the leading equipment used when providing advanced cardiac care; and,

**WHEREAS**, Stryker/Physio Control is the manufacturer of the Lifepak 15. Consequently, to maintain compatibility with the remaining units that are in-service and to receive appropriate trade-in credit for older units, the vendor is considered the only known source for this purchase per the Procurement Ordinance, Article 3-114.

**NOW, THEREFORE, THE COUNCIL OF COLUMBUS, GEORGIA, HEREBY RESOLVES AS FOLLOWS:**

That the City Manager is hereby authorized to purchase one (1) Stryker Lifepak 15 Monitor Defibrillator and accessories from Stryker Medical (Chicago, IL) in the amount of \$45,346.29. Funds are budgeted in the FY24 Budget: General Fund - Fire & EMS – Fire & EMS Operations – Capital Expenditure Over \$5,000; 0101 – 410 – 2100 – FOPR – 7761.

Introduced at a regular meeting of the Council of Columbus, Georgia, held the \_\_\_\_\_ day of \_\_\_\_\_, 2023 and adopted at said meeting by the affirmative vote of \_\_\_\_\_ members of said Council.

- Councilor Allen voting \_\_\_\_\_
- Councilor Barnes voting \_\_\_\_\_
- Councilor Begly voting \_\_\_\_\_
- Councilor Cogle voting \_\_\_\_\_
- Councilor Crabb voting \_\_\_\_\_
- Councilor Davis voting \_\_\_\_\_
- Councilor Garrett voting \_\_\_\_\_
- Councilor Huff voting \_\_\_\_\_
- Councilor Thomas voting \_\_\_\_\_
- Councilor Tucker voting \_\_\_\_\_

\_\_\_\_\_  
Sandra T. Davis, Clerk of Council

\_\_\_\_\_  
B.H. "Skip" Henderson III, Mayor